

Manual Supplement

| | | | |
|----------------|-----------------------------|-------------------|------|
| Manual Title: | 902 Calibration Information | Supplement Issue: | 1 |
| Part Number: | 2649733 | Issue Date: | 8/13 |
| Print Date: | July 2006 | Page Count: | 1 |
| Revision/Date: | | | |

This supplement contains information necessary to ensure the accuracy of the above manual.



Change #1

On page 2, add:

- **Measure a known voltage to verify meter operation.**

Change #2, 66422, 66561

On page 3, add the following to the **Symbols** table:

| | |
|---|--|
| CAT II | Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation. |
| CAT III | Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation. |
| CAT IV | Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation. |
|  | Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1. |
|  | Conforms to relevant South Korean EMC Standards. |

On page 4, remove **EMI, RFI, EMC, Design Standards and Compliance, Agency Approvals** and **Measurement Category** and replace **IP Rating** with:

IP Rating IEC 60529: IP40

Safety IEC 61010-1: 600 V CAT III, Pollution Degree 2

Electromagnetic Environment IEC 61326-1: Portable

Electromagnetic Compatibility Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.